
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 003633
8:08 10/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:12
PRESSURIZING 4:13
STERILIZING 5:10
136.0 C 220kPa 5:10
136.1 C 221kPa 5:40
136.1 C 218kPa 6:10
136.4 C 223kPa 6:40
136.3 C 220kPa 7:10
136.2 C 222kPa 7:40
136.4 C 223kPa 8:10
136.2 C 221kPa 8:40
136.2 C 221kPa 8:40
VENTING 8:40
AIR DRYING 9:31
CYCLE COMPLETE 9:50

Digital Signature #
03C4EEE10F788BE0
