

-----  
Midmark M11 - v1.0.0  
Total Cycles: 977

Name: RITTER 1  
Sterilizer ID: V2178198  
07/07/2021 02:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 213.1 F 0.0  
2:00 210.0 F 3.7  
4:00 218.9 F 8.9  
6:00 230.2 F 8.5  
8:00 233.1 F 10.0  
10:00 238.3 F 8.5  
12:00 245.6 F 13.5  
14:00 244.5 F 12.4  
16:00 257.1 F 19.0  
18:00 268.1 F 26.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.2  
1:00 272.7 F 28.9  
2:00 271.6 F 28.1  
3:00 272.8 F 29.0  
4:00 272.1 F 28.4  
  
Min 271.5 F 28.1  
Max 272.8 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 215.6 F 0.0  
5:00 230.3 F 0.0  
10:00 230.9 F 0.0  
15:00 226.8 F 0.0  
20:00 225.0 F 0.0  
25:00 226.8 F 0.0  
30:00 229.1 F 0.0  
  
Drying Complete

FILLING: 2:17  
HEATING: 19:16  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:47  
CYCLE COMPLETE

07/07/2021 03:08 PM  
ACCEPTED: X  
REJECTED:  
Operator: U02