
Midmark M11 - v1.0.5
Total Cycles: 2279

Name: RITTER2
Sterilizer ID: V2313228
04/20/2023 03:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.0 F 0.0
2:00 196.6 F 3.9
4:00 207.7 F 9.3
6:00 223.2 F 7.8
8:00 233.9 F 10.7
10:00 234.1 F 8.5
12:00 240.1 F 10.4
14:00 248.6 F 14.8
16:00 261.3 F 21.9
18:00 272.5 F 29.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.6 F 29.3
2:00 271.5 F 28.9
3:00 272.6 F 29.4
4:00 271.6 F 28.7

Min 271.4 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.6 F 0.0
5:00 208.1 F 0.0
10:00 218.4 F 0.0
15:00 218.5 F 0.0
20:00 215.5 F 0.0
25:00 211.4 F 0.0
30:00 207.9 F 0.0

Drying Complete

FILLING: 2:33
HEATING: 18:25
STERILIZING: 4:00
VENTING: 3:20
DRYING: 30:00
TOTAL CYCLE: 00:58:17
CYCLE COMPLETE

04/20/2023 04:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____