
Midmark M11 - v1.0.0
Total Cycles: 2649

Name: RITTER 1
Sterilizer ID: V2178198
12/06/2022 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 108.5 F 0.0
2:00 133.7 F 0.0
4:00 153.6 F 1.1
6:00 169.5 F 3.7
8:00 192.7 F 8.0
10:00 211.4 F 12.6
12:00 220.9 F 6.4
14:00 233.3 F 11.1
16:00 231.9 F 7.5
18:00 244.9 F 12.9
20:00 243.7 F 12.0
22:00 256.3 F 18.7
24:00 267.6 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.3 F 28.6
2:00 272.4 F 29.1
3:00 271.8 F 28.2
4:00 272.6 F 29.0

Min 271.5 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 210.5 F 0.0
10:00 221.0 F 0.0
15:00 222.0 F 0.0
20:00 230.1 F 0.0
25:00 234.8 F 0.0
30:00 221.2 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 25:19
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:03:24
CYCLE COMPLETE

12/06/2022 04:57 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____