

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2439

Name: RITTER2  
Sterilizer ID: V2313228  
05/30/2023 04:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 198.7 F 0.0  
2:00 188.7 F 3.6  
4:00 216.2 F 8.7  
6:00 224.8 F 10.1  
8:00 231.4 F 9.1  
10:00 238.8 F 10.7  
12:00 243.6 F 12.7  
14:00 243.1 F 12.0  
16:00 255.9 F 18.7  
18:00 267.2 F 25.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.7 F 29.4  
2:00 271.8 F 28.7  
3:00 271.8 F 29.3  
4:00 272.7 F 29.3  
  
Min 271.1 F 28.5  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 192.3 F 0.0  
5:00 234.3 F 0.0  
10:00 231.0 F 0.0  
15:00 222.1 F 0.0  
20:00 217.8 F 0.0  
25:00 218.9 F 0.0  
30:00 215.3 F 0.0  
  
Drying Complete

FILLING: 2:32  
HEATING: 19:28  
STERILIZING: 4:00  
VENTING: 3:29  
DRYING: 30:00  
TOTAL CYCLE: 00:59:29  
CYCLE COMPLETE

05/30/2023 04:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_