

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2090

Name: RITTER2  
Sterilizer ID: V2313228  
02/27/2023 08:16 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.5 F 0.0  
2:00 79.3 F 0.0  
4:00 91.8 F 0.0  
6:00 114.6 F 0.1  
8:00 156.1 F 0.4  
10:00 184.7 F 3.2  
12:00 198.9 F 7.0  
14:00 213.0 F 11.0  
16:00 219.9 F 5.6  
18:00 232.0 F 9.6  
20:00 237.6 F 10.6  
22:00 242.4 F 12.2  
24:00 241.0 F 11.0  
26:00 253.6 F 17.4  
28:00 264.8 F 24.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.1 F 28.7  
2:00 272.7 F 29.3  
3:00 272.0 F 29.1  
4:00 271.9 F 28.6  
  
Min 271.5 F 28.5  
Max 272.7 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 171.8 F 0.0  
5:00 208.2 F 0.0  
10:00 216.8 F 0.0  
15:00 215.8 F 0.0  
20:00 218.1 F 0.0  
25:00 216.7 F 0.0  
30:00 214.5 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 29:51  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 01:07:47  
CYCLE COMPLETE

02/27/2023 09:23 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_