
Midmark M11 - v1.0.0
Total Cycles: 2920

Name: RITTER 1
Sterilizer ID: V2178198
02/21/2023 08:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 75.3 F 0.0
2:00 75.7 F 0.0
4:00 85.2 F 0.1
6:00 122.9 F 0.3
8:00 155.8 F 1.5
10:00 177.0 F 4.5
12:00 193.7 F 7.7
14:00 205.8 F 11.4
16:00 218.7 F 4.5
18:00 228.6 F 8.9
20:00 237.5 F 13.3
22:00 237.1 F 9.9
24:00 237.5 F 9.0
26:00 247.0 F 13.7
28:00 258.4 F 19.9
30:00 268.8 F 26.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.5 F 28.9
2:00 271.4 F 28.4
3:00 272.6 F 28.9
4:00 271.3 F 28.2

Min 271.3 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 219.4 F 0.0
10:00 184.6 F 0.0
15:00 187.3 F 0.0
20:00 188.1 F 0.0
25:00 189.2 F 0.0
30:00 191.2 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 31:10
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 01:09:12
CYCLE COMPLETE

02/21/2023 09:38 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____