
Midmark M11 - v1.0.5
Total Cycles: 1626

Name: RITTER2
Sterilizer ID: V2313228
10/19/2022 02:48 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.8 F 0.0
2:00 175.9 F 3.4
4:00 184.7 F 7.7
6:00 210.8 F 12.8
8:00 222.3 F 7.7
10:00 234.8 F 13.3
12:00 237.8 F 11.0
14:00 237.2 F 9.8
16:00 251.1 F 16.6
18:00 262.9 F 23.5
20:00 272.4 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.3
1:00 272.8 F 30.0
2:00 271.7 F 29.2
3:00 271.8 F 30.0
4:00 272.5 F 29.8

Min 271.2 F 29.2
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.5 F 0.0
5:00 215.2 F 0.0
10:00 236.0 F 0.0
15:00 234.4 F 0.0
20:00 227.0 F 0.0
25:00 226.8 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 1:50
HEATING: 20:15
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 00:59:21
CYCLE COMPLETE

10/19/2022 03:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____