
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 007815
9:03 03/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:21
PRESSURIZING 3:26
STERILIZING 4:30
136.4 C 221kPa 4:30
136.1 C 221kPa 5:00
CYCLE ABORTED 5:19
STOP BUTTON PRESSED
NOT STERILE
END TIME 5:19

Digital Signature #
ADEDEBC367B2EF48
