

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 005122  
14:06 14/02/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:36  
PRESSURIZING 2:40  
STERILIZING 3:37  
136.4 C 221kPa 3:37  
136.4 C 225kPa 4:07  
136.4 C 225kPa 4:37  
136.4 C 225kPa 5:07  
136.4 C 225kPa 5:37  
136.2 C 221kPa 6:07  
136.4 C 224kPa 6:37  
136.3 C 224kPa 7:07  
136.3 C 224kPa 7:07  
VENTING 7:08  
AIR DRYING 7:57  
CYCLE COMPLETE 11:55

-----  
Digital Signature #  
6D49BF67A1B7F83E  
-----