
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001632
12:22 29/11/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:58
PRESSURIZING 2:04
STERILIZING 3:03
136.0 C 222kPa 3:03
136.2 C 224kPa 3:33
136.4 C 226kPa 4:03
136.3 C 223kPa 4:33
136.3 C 225kPa 5:03
136.4 C 225kPa 5:33
136.3 C 225kPa 6:03
136.4 C 226kPa 6:33
136.4 C 226kPa 6:33
VENTING 6:34
AIR DRYING 7:23
CYCLE COMPLETE 7:30

Digital Signature #
610DBD4D47ED719E
