
Midmark M11 - v1.0.0
Total Cycles: 2755

Name: RITTER 1
Sterilizer ID: V2178198
01/10/2023 08:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.2 F 0.0
2:00 82.1 F 0.0
4:00 95.4 F 0.0
6:00 116.3 F 0.0
8:00 148.9 F 0.0
10:00 177.3 F 0.0
12:00 202.1 F 3.4
14:00 216.5 F 6.5
16:00 224.2 F 9.7
18:00 232.2 F 13.4
20:00 234.1 F 8.9
22:00 243.1 F 12.3
24:00 242.9 F 11.5
26:00 239.3 F 9.8
28:00 252.1 F 16.1
30:00 262.9 F 22.4
32:00 272.8 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.1
1:00 272.7 F 28.8
2:00 271.8 F 28.0
3:00 272.0 F 28.6
4:00 272.5 F 28.6

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.4 F 0.0
5:00 203.7 F 0.0
10:00 226.5 F 0.0
15:00 228.9 F 0.0
20:00 235.4 F 0.0
25:00 220.8 F 0.0
30:00 217.8 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 32:23
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:10:19
CYCLE COMPLETE

01/10/2023 09:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____