
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 001013
14:49 16/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:06
STERILIZING 3:01
136.0 C 230kPa 3:01
136.0 C 231kPa 3:31
136.2 C 232kPa 4:01
136.4 C 233kPa 4:31
136.2 C 233kPa 5:01
136.3 C 232kPa 5:31
136.4 C 233kPa 6:01
136.1 C 232kPa 6:31
136.1 C 232kPa 6:31
VENTING 6:31
AIR DRYING 7:23
CYCLE COMPLETE 7:28

Digital Signature #
AFCA36C87F2A6161
