
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
1.2uS / 0.7ppm
CYCLE NUMBER 000628
14:38 02/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:22
STERILIZING 4:37
136.0 C 220kPa 4:37
136.4 C 231kPa 5:07
136.4 C 232kPa 5:37
136.4 C 233kPa 6:07
136.4 C 231kPa 6:37
136.4 C 233kPa 7:07
136.4 C 233kPa 7:37
136.4 C 233kPa 8:07
136.4 C 233kPa 8:07
VENTING 8:07
AIR DRYING 9:02
CYCLE COMPLETE 24:02

Digital Signature #
7050C469D20CCCD9
