
Midmark M11 - v1.0.5
Total Cycles: 1476

Name: RITTER2
Sterilizer ID: V2313228
09/12/2022 12:49 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.1 F 0.0
2:00 163.7 F 3.0
4:00 176.5 F 6.9
6:00 210.2 F 11.5
8:00 222.3 F 6.4
10:00 233.8 F 11.7
12:00 233.5 F 8.8
14:00 240.0 F 10.6
16:00 245.3 F 13.5
18:00 257.5 F 20.0
20:00 268.4 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.6
1:00 272.6 F 29.7
2:00 272.7 F 30.1
3:00 271.9 F 30.0
4:00 272.4 F 29.7

Min 271.8 F 29.5
Max 272.8 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.9 F 0.0
5:00 233.8 F 0.0
10:00 234.3 F 0.0
15:00 227.0 F 0.0
20:00 222.7 F 0.0
25:00 226.6 F 0.0
30:00 226.6 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 21:58
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:00:09
CYCLE COMPLETE

09/12/2022 01:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____