
Midmark M11 - v1.0.5
Total Cycles: 1869

Name: RITTER2
Sterilizer ID: V2313228
12/21/2022 08:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 78.3 F 0.0
2:00 79.0 F 0.0
4:00 85.5 F 0.0
6:00 103.5 F 0.0
8:00 164.4 F 0.0
10:00 190.5 F 1.9
12:00 202.7 F 5.5
14:00 217.8 F 9.0
16:00 226.3 F 13.2
18:00 230.3 F 8.8
20:00 239.0 F 12.0
22:00 241.0 F 11.9
24:00 239.6 F 10.7
26:00 252.5 F 17.1
28:00 263.7 F 23.8
30:00 271.9 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.4 F 29.6
2:00 271.5 F 29.0
3:00 272.4 F 29.7
4:00 271.7 F 29.0

Min 271.4 F 28.9
Max 272.4 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.4 F 0.0
5:00 171.5 F 0.0
10:00 196.6 F 0.0
15:00 206.1 F 0.0
20:00 207.5 F 0.0
25:00 209.5 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 30:07
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:08:05
CYCLE COMPLETE

12/21/2022 10:03 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____