

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5460

Name: RITTER 1  
Sterilizer ID: V2178198  
03/28/2025 09:58 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 175.0 F 0.0  
2:00 169.0 F 2.6  
4:00 177.0 F 6.3  
6:00 193.7 F 10.4  
8:00 213.3 F 4.2  
10:00 219.4 F 7.1  
12:00 231.7 F 11.4  
14:00 232.7 F 7.4  
16:00 244.2 F 12.9  
18:00 242.2 F 11.2  
20:00 254.1 F 17.2  
22:00 264.6 F 23.5  
24:00 271.0 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.6  
1:00 272.7 F 29.2  
2:00 272.0 F 28.6  
3:00 272.0 F 29.1  
4:00 272.5 F 29.1

Min 271.4 F 28.4  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 201.7 F 0.0  
5:00 214.3 F 0.0  
10:00 223.6 F 0.0  
15:00 227.0 F 0.0  
20:00 226.6 F 0.0  
25:00 234.0 F 0.0  
30:00 231.6 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 24:51  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:03:51  
CYCLE COMPLETE

03/28/2025 11:01 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_