
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000931
11:23 08/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:14
STERILIZING 4:15
136.0 C 220kPa 4:15
136.4 C 225kPa 4:45
136.2 C 223kPa 5:15
135.9 C 220kPa 5:45
136.4 C 225kPa 6:15
136.2 C 223kPa 6:45
136.2 C 220kPa 7:15
136.3 C 224kPa 7:45
136.3 C 224kPa 7:45
VENTING 7:46
AIR DRYING 8:41
CYCLE COMPLETE 8:54

Digital Signature #
6F254D4677EFD1F4
