
Midmark M11 - v1.0.0
Total Cycles: 3646

Name: RITTER 1
Sterilizer ID: V2178198
08/14/2023 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 173.0 F 0.0
2:00 170.4 F 2.9
4:00 179.1 F 6.5
6:00 203.0 F 11.1
8:00 217.3 F 4.4
10:00 229.2 F 9.1
12:00 237.0 F 11.4
14:00 239.9 F 11.2
16:00 237.9 F 9.4
18:00 251.0 F 15.7
20:00 261.9 F 22.0
22:00 271.9 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 271.9 F 28.5
2:00 272.5 F 29.0
3:00 271.2 F 28.5
4:00 272.0 F 28.5

Min 271.2 F 28.2
Max 272.6 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.9 F 0.0
5:00 201.8 F 0.0
10:00 197.7 F 0.0
15:00 183.5 F 0.0
20:00 190.6 F 0.0
25:00 185.7 F 0.0
30:00 181.8 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 22:29
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:01:48
CYCLE COMPLETE

08/14/2023 04:05 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____