
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002239
11:51 11/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:01
PRESSURIZING 2:06
STERILIZING 3:05
136.0 C 220kPa 3:05
136.4 C 222kPa 3:35
136.3 C 224kPa 4:05
136.4 C 223kPa 4:35
136.4 C 226kPa 5:05
136.3 C 225kPa 5:35
136.4 C 223kPa 6:05
136.0 C 222kPa 6:35
136.0 C 222kPa 6:35
VENTING 6:35
AIR DRYING 7:30
CYCLE COMPLETE 15:27

Digital Signature #
171BC4B9DC8BA81C
