
Midmark M11 - v1.0.5
Total Cycles: 1492

Name: RITTER2
Sterilizer ID: V2313228
09/14/2022 02:17 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.3 F 0.0
2:00 198.0 F 4.1
4:00 209.9 F 10.5
6:00 225.2 F 6.9
8:00 240.2 F 13.3
10:00 242.3 F 12.1
12:00 243.0 F 11.9
14:00 257.5 F 19.6
16:00 270.0 F 27.5

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.6
1:00 272.4 F 28.9
2:00 272.6 F 29.4
3:00 271.8 F 28.5
4:00 272.8 F 29.3

Min 271.6 F 28.4
Max 272.9 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.3 F 0.0
5:00 214.1 F 0.0
10:00 226.6 F 0.0
15:00 232.2 F 0.0
20:00 224.2 F 0.0
25:00 220.4 F 0.0
30:00 216.0 F 0.0
Drying Complete

FILLING: 2:34
HEATING: 16:49
STERILIZING: 4:00
VENTING: 2:59
DRYING: 30:00
TOTAL CYCLE: 00:56:22
CYCLE COMPLETE

09/14/2022 03:13 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____