
Midmark M11 - v1.0.0
Total Cycles: 1896

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2022 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.4 F 0.0
2:00 190.3 F 3.2
4:00 207.2 F 8.0
6:00 218.2 F 9.8
8:00 230.8 F 9.1
10:00 239.5 F 9.7
12:00 242.9 F 12.5
14:00 241.9 F 11.1
16:00 254.8 F 17.9
18:00 266.4 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.5 F 28.3
3:00 272.5 F 29.2
4:00 272.0 F 28.6

Min 271.2 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.9 F 0.0
5:00 233.8 F 0.0
10:00 228.3 F 0.0
15:00 225.4 F 0.0
20:00 222.2 F 0.0
25:00 223.8 F 0.0
30:00 225.0 F 0.0

Drying Complete

FILLING: 3:42
HEATING: 19:35
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 01:00:27
CYCLE COMPLETE

05/02/2022 03:34 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____