
Midmark M11 - v1.0.5
Total Cycles: 1401

Name: RITTER2
Sterilizer ID: V2313228
08/16/2022 02:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.1 F 0.0
2:00 154.9 F 0.0
4:00 163.6 F 0.5
6:00 195.4 F 5.0
8:00 209.4 F 10.0
10:00 221.9 F 6.9
12:00 234.7 F 11.0
14:00 234.8 F 8.8
16:00 240.6 F 10.2
18:00 248.5 F 14.8
20:00 260.8 F 21.8
22:00 271.9 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 272.8 F 29.5
2:00 271.7 F 29.0
3:00 272.8 F 29.6
4:00 271.9 F 28.9

Min 271.7 F 28.8
Max 272.9 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 203.7 F 0.0
10:00 231.9 F 0.0
15:00 227.4 F 0.0
20:00 225.9 F 0.0
25:00 227.0 F 0.0
30:00 229.2 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 22:33
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:01:34
CYCLE COMPLETE

08/16/2022 03:36 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____