
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002108
12:06 02/10/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:16
PRESSURIZING 2:21
STERILIZING 3:25
136.0 C 220kPa 3:25
136.1 C 223kPa 3:55
136.4 C 222kPa 4:25
136.4 C 226kPa 4:55
136.3 C 224kPa 5:25
136.3 C 225kPa 5:55
136.3 C 224kPa 6:25
136.4 C 222kPa 6:55
136.4 C 222kPa 6:55
VENTING 6:55
AIR DRYING 7:46
CYCLE COMPLETE 37:46

Digital Signature #
032432EDF732438E
