
Midmark M11 - v1.0.0
Total Cycles: 3532

Name: RITTER 1
Sterilizer ID: V2178198
07/17/2023 09:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 133.2 F 0.0
2:00 141.0 F 0.0
4:00 143.3 F 0.5
6:00 166.8 F 2.6
8:00 189.7 F 5.4
10:00 209.6 F 8.9
12:00 219.1 F 13.0
14:00 226.4 F 7.3
16:00 236.6 F 11.7
18:00 233.4 F 8.4
20:00 243.8 F 12.3
22:00 243.8 F 12.3
24:00 255.4 F 18.2
26:00 265.6 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.3
1:00 272.6 F 29.1
2:00 271.9 F 28.4
3:00 271.5 F 28.6
4:00 272.8 F 29.1

Min 271.2 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.9 F 0.0
5:00 203.0 F 0.0
10:00 193.0 F 0.0
15:00 191.8 F 0.0
20:00 195.8 F 0.0
25:00 195.0 F 0.0
30:00 196.2 F 0.0
Drying Complete

FILLING: 0:47
HEATING: 27:55
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:06:14
CYCLE COMPLETE

07/17/2023 10:25 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____