

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4983

Name: RITTER 1  
Sterilizer ID: V2178198  
10/10/2024 08:11 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 71.3 F 0.0  
2:00 78.2 F 0.0  
4:00 88.4 F 0.0  
6:00 120.8 F 0.1  
8:00 149.9 F 0.4  
10:00 178.2 F 1.0  
12:00 199.7 F 4.2  
14:00 210.3 F 7.4  
16:00 222.1 F 11.0  
18:00 224.7 F 5.9  
20:00 234.7 F 10.1  
22:00 235.4 F 8.3  
24:00 244.7 F 12.7  
26:00 242.0 F 11.0  
28:00 254.8 F 17.5  
30:00 265.9 F 24.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.0  
1:00 272.6 F 28.6  
2:00 271.5 F 28.0  
3:00 272.6 F 28.6  
4:00 271.5 F 27.8

Min 271.4 F 27.8  
Max 272.7 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.7 F 0.0  
5:00 202.8 F 0.0  
10:00 191.6 F 0.0  
15:00 214.6 F 0.0  
20:00 218.7 F 0.0  
25:00 215.9 F 0.0  
30:00 215.4 F 0.0  
Drying Complete

FILLING: 0:43  
HEATING: 31:42  
STERILIZING: 4:00  
VENTING: 3:23  
DRYING: 30:00  
TOTAL CYCLE: 01:09:48  
CYCLE COMPLETE

10/10/2024 09:20 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_