

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1148

Name: RITTER2  
Sterilizer ID: V2313228  
05/31/2022 02:23 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 181.8 F 0.0  
2:00 175.0 F 3.4  
4:00 186.5 F 8.2  
6:00 207.8 F 11.3  
8:00 228.1 F 8.1  
10:00 239.4 F 13.3  
12:00 239.0 F 11.0  
14:00 237.3 F 9.6  
16:00 251.4 F 16.4  
18:00 263.4 F 23.3

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.9  
1:00 272.1 F 29.5  
2:00 271.6 F 28.7  
3:00 272.3 F 29.6  
4:00 271.7 F 29.0  
  
Min 271.0 F 28.7  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.4 F 0.0  
5:00 232.7 F 0.0  
10:00 223.3 F 0.0  
15:00 228.6 F 0.0  
20:00 224.9 F 0.0  
25:00 226.6 F 0.0  
30:00 199.0 F 0.0  
Drying Complete

FILLING: 1:55  
HEATING: 19:58  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:03  
CYCLE COMPLETE

05/31/2022 03:22 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_