
Midmark M11 - v1.0.0
Total Cycles: 1911

Name: RITTER 1
Sterilizer ID: V2178198
05/05/2022 07:03 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.3 F 0.0
2:00 94.2 F 0.0
4:00 91.3 F 0.0
6:00 126.7 F 1.3
8:00 160.1 F 3.3
10:00 186.4 F 6.5
12:00 209.5 F 10.5
14:00 217.6 F 5.6
16:00 229.1 F 7.3
18:00 237.3 F 11.5
20:00 232.5 F 7.3
22:00 245.5 F 13.2
24:00 242.6 F 11.3
26:00 255.7 F 18.0
28:00 267.1 F 25.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 27.9
1:00 272.5 F 28.4
2:00 272.0 F 28.3
3:00 272.2 F 28.2
4:00 272.3 F 28.6

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.6 F 0.0
5:00 168.9 F 0.0
10:00 164.9 F 0.0
15:00 188.8 F 0.0
20:00 173.2 F 0.0
25:00 174.0 F 0.0
30:00 164.5 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:27
STERILIZING: 4:00
VENTING: 2:50
DRYING: 30:00
TOTAL CYCLE: 01:07:02
CYCLE COMPLETE

05/05/2022 08:55 AM
ACCEPTED: X
REJECTED:
Operator: U02