

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4236

Name: RITTER 1  
Sterilizer ID: V2178198  
01/26/2024 12:03 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 182.5 F 0.0  
2:00 178.4 F 2.8  
4:00 191.7 F 7.0  
6:00 212.3 F 11.7  
8:00 226.4 F 5.6  
10:00 230.4 F 10.3  
12:00 235.3 F 8.4  
14:00 244.1 F 12.5  
16:00 241.7 F 11.0  
18:00 253.9 F 17.1  
20:00 264.6 F 23.5  
22:00 270.9 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.5  
1:00 272.6 F 28.7  
2:00 272.7 F 29.2  
3:00 271.6 F 28.4  
4:00 272.5 F 29.0

Min 271.5 F 28.4  
Max 272.9 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 196.0 F 0.0  
5:00 242.3 F 0.0  
10:00 219.8 F 0.0  
15:00 227.4 F 0.0  
20:00 210.5 F 0.0  
25:00 214.7 F 0.0  
30:00 234.8 F 0.0  
Drying Complete

FILLING: 1:46  
HEATING: 22:44  
STERILIZING: 4:00  
VENTING: 3:47  
DRYING: 30:00  
TOTAL CYCLE: 01:02:17  
CYCLE COMPLETE

01/26/2024 01:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_