
Midmark M11 - v1.0.0
Total Cycles: 2594

Name: RITTER 1
Sterilizer ID: V2178198
11/22/2022 11:50 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.4 F 0.0
2:00 188.3 F 4.0
4:00 194.0 F 9.1
6:00 217.6 F 7.5
8:00 228.6 F 9.0
10:00 235.6 F 9.8
12:00 243.3 F 12.0
14:00 240.4 F 10.4
16:00 254.4 F 17.3
18:00 265.9 F 24.3
20:00 272.8 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.4
1:00 272.7 F 29.1
2:00 271.7 F 28.3
3:00 272.7 F 29.1
4:00 271.9 F 28.4

Min 271.3 F 28.3
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.9 F 0.0
5:00 241.4 F 0.0
10:00 228.1 F 0.0
15:00 226.8 F 0.0
20:00 226.2 F 0.0
Dry Cycle Aborted

FILLING: 1:49
HEATING: 20:14
STERILIZING: 4:00
VENTING: 3:12
DRYING: 24:34
TOTAL CYCLE: 00:53:49

11/22/2022 12:43 PM
ACCEPTED: X
REJECTED:
Operator: U02