
Midmark M11 - v1.0.5
Total Cycles: 1372

Name: RITTER2
Sterilizer ID: V2313228
08/08/2022 07:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.0 F 0.0
2:00 75.0 F 0.0
4:00 100.7 F 0.0
6:00 120.9 F 0.0
8:00 139.5 F 0.3
10:00 167.4 F 2.7
12:00 187.0 F 5.9
14:00 212.4 F 9.6
16:00 222.4 F 11.3
18:00 226.9 F 8.0
20:00 236.7 F 12.3
22:00 235.0 F 9.3
24:00 239.7 F 10.4
26:00 245.3 F 13.4
28:00 256.5 F 19.4
30:00 266.7 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.0
1:00 272.3 F 29.5
2:00 271.8 F 29.5
3:00 271.9 F 29.2
4:00 272.0 F 29.7

Min 271.4 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 204.4 F 0.0
10:00 211.7 F 0.0
15:00 213.6 F 0.0
20:00 215.8 F 0.0
25:00 216.6 F 0.0
30:00 218.3 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 31:35
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:31
CYCLE COMPLETE

08/08/2022 08:31 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____