
Midmark M11 - v1.0.5
Total Cycles: 1487

Name: RITTER2
Sterilizer ID: V2313228
09/14/2022 07:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.9 F 0.0
2:00 75.8 F 0.0
4:00 96.8 F 0.0
6:00 100.8 F 0.5
8:00 147.5 F 3.4
10:00 164.4 F 6.9
12:00 200.8 F 10.7
14:00 215.6 F 6.2
16:00 219.3 F 8.0
18:00 231.3 F 12.1
20:00 230.8 F 8.2
22:00 242.8 F 13.5
24:00 241.5 F 11.8
26:00 253.6 F 18.0
28:00 264.3 F 24.5
30:00 271.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.3
1:00 272.7 F 30.0
2:00 271.6 F 29.2
3:00 272.4 F 30.2
4:00 272.3 F 29.6

Min 271.4 F 29.1
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.3 F 0.0
5:00 208.6 F 0.0
10:00 221.2 F 0.0
15:00 225.5 F 0.0
20:00 220.0 F 0.0
25:00 219.1 F 0.0
30:00 220.9 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 30:04
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:08:01
CYCLE COMPLETE

09/14/2022 08:32 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____