
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004143
8:39 04/08/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:15
STERILIZING 4:23
136.4 C 224kPa 4:23
136.1 C 222kPa 4:53
136.1 C 221kPa 5:23
136.3 C 222kPa 5:53
136.0 C 222kPa 6:23
136.3 C 224kPa 6:53
136.3 C 222kPa 7:23
136.3 C 221kPa 7:53
136.3 C 221kPa 7:53
VENTING 7:53
AIR DRYING 8:44
CYCLE COMPLETE 8:45

Digital Signature #
026B4E19361F402E
