
Midmark M11 - v1.0.0
Total Cycles: 2712

Name: RITTER 1
Sterilizer ID: V2178198
12/21/2022 03:35 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.3 F 0.0
2:00 209.5 F 4.4
4:00 210.7 F 9.2
6:00 222.6 F 7.7
8:00 231.7 F 9.6
10:00 239.0 F 9.0
12:00 243.4 F 12.6
14:00 242.1 F 11.2
16:00 255.2 F 17.9
18:00 266.5 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.3
1:00 272.6 F 28.9
2:00 271.6 F 28.2
3:00 272.7 F 29.0
4:00 271.9 F 28.3

Min 271.5 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.7 F 0.0
5:00 225.2 F 0.0
10:00 239.2 F 0.0
15:00 230.2 F 0.0
20:00 228.6 F 0.0
25:00 225.5 F 0.0
30:00 224.1 F 0.0
Drying Complete

FILLING: 2:11
HEATING: 19:34
STERILIZING: 4:00
VENTING: 3:12
DRYING: 30:00
TOTAL CYCLE: 00:58:57
CYCLE COMPLETE

12/21/2022 04:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____