
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 002243
8:49 12/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:24
PRESSURIZING 2:30
STERILIZING 3:40
136.4 C 220kPa 3:40
136.2 C 222kPa 4:10
136.4 C 221kPa 4:40
136.3 C 222kPa 5:10
136.4 C 221kPa 5:40
136.1 C 221kPa 6:10
136.4 C 223kPa 6:40
136.3 C 223kPa 7:10
136.3 C 223kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 38:02

Digital Signature #
3CA3883B26BDA67E
