
Midmark M11 - v1.0.5
Total Cycles: 1331

Name: RITTER2
Sterilizer ID: V2313228
07/21/2022 11:29 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 190.6 F 0.0
2:00 183.7 F 1.2
4:00 216.2 F 6.6
6:00 231.2 F 12.6
8:00 233.3 F 9.0
10:00 239.1 F 10.2
12:00 244.3 F 13.2
14:00 243.3 F 12.6
16:00 256.5 F 19.4
18:00 267.7 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.1 F 29.5
2:00 271.1 F 29.2
3:00 272.3 F 29.5
4:00 272.3 F 29.5

Min 271.0 F 28.9
Max 273.1 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 231.0 F 0.0
10:00 199.8 F 0.0
15:00 177.1 F 0.0
Dry Cycle Aborted

FILLING: 2:15
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:14
DRYING: 17:32
TOTAL CYCLE: 00:46:21

07/21/2022 12:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____