
Midmark M11 - v1.0.5
Total Cycles: 2062

Name: RITTER2
Sterilizer ID: V2313228
02/16/2023 11:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 178.9 F 0.0
2:00 176.5 F 0.2
4:00 187.0 F 3.7
6:00 205.7 F 8.2
8:00 221.9 F 13.3
10:00 228.9 F 9.6
12:00 235.5 F 8.9
14:00 243.7 F 12.2
16:00 245.7 F 13.2
18:00 258.0 F 20.2
20:00 269.1 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.8 F 29.8
2:00 272.2 F 29.2
3:00 271.3 F 29.1
4:00 272.8 F 29.8

Min 271.2 F 28.8
Max 272.9 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.3 F 0.0
5:00 222.1 F 0.0
10:00 231.2 F 0.0
15:00 228.7 F 0.0
20:00 227.8 F 0.0
25:00 229.6 F 0.0
30:00 229.5 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 21:06
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:59:11
CYCLE COMPLETE

02/16/2023 12:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____