

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001228  
10:42 26/08/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:02  
PRESSURIZING 3:16  
STERILIZING 4:22  
136.0 C 220kPa 4:22  
136.4 C 225kPa 4:52  
136.4 C 225kPa 5:22  
136.4 C 225kPa 5:52  
136.3 C 221kPa 6:22  
136.4 C 221kPa 6:52  
136.4 C 225kPa 7:22  
136.4 C 224kPa 7:52  
136.4 C 224kPa 7:52  
VENTING 7:52  
AIR DRYING 9:14  
CYCLE COMPLETE 24:14

-----  
Digital Signature #  
9D510C2738B3871C  
-----