
Midmark M11 - v1.0.5
Total Cycles: 1783

Name: RITTER2
Sterilizer ID: V2313228
11/29/2022 10:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.6 F 0.0
2:00 176.3 F 0.2
4:00 185.1 F 1.5
6:00 207.6 F 6.5
8:00 225.2 F 11.8
10:00 230.3 F 8.6
12:00 238.5 F 10.2
14:00 244.5 F 12.6
16:00 246.0 F 13.4
18:00 259.1 F 20.6
20:00 270.6 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.7
1:00 272.9 F 29.3
2:00 271.6 F 28.8
3:00 272.6 F 29.4
4:00 271.8 F 28.7

Min 271.5 F 28.6
Max 272.9 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 219.9 F 0.0
10:00 236.3 F 0.0
15:00 229.2 F 0.0
20:00 226.1 F 0.0
25:00 224.7 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 1:18
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 00:59:12
CYCLE COMPLETE

11/29/2022 11:18 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____