
Midmark M11 - v1.0.0
Total Cycles: 1805

Name: RITTER 1
Sterilizer ID: V2178198
04/07/2022 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.8 F 0.0
2:00 111.6 F 0.0
4:00 136.3 F 0.9
6:00 165.8 F 4.3
8:00 189.3 F 8.4
10:00 207.9 F 12.6
12:00 220.1 F 5.4
14:00 230.5 F 9.8
16:00 235.4 F 9.2
18:00 241.3 F 11.5
20:00 238.1 F 9.5
22:00 251.5 F 16.0
24:00 263.0 F 22.6
26:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 271.7 F 28.3
3:00 272.3 F 29.1
4:00 272.6 F 28.8

Min 271.3 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.3 F 0.0
5:00 230.9 F 0.0
10:00 220.2 F 0.0
15:00 234.7 F 0.0
20:00 228.7 F 0.0
25:00 232.4 F 0.0
30:00 225.3 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 26:17
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 01:04:19
CYCLE COMPLETE

04/07/2022 02:27 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____