
Midmark M11 - v1.0.0
Total Cycles: 2361

Name: RITTER 1
Sterilizer ID: V2178198
09/22/2022 01:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.6 F 0.0
2:00 200.4 F 3.9
4:00 207.6 F 8.4
6:00 222.0 F 12.5
8:00 227.5 F 8.7
10:00 238.8 F 10.7
12:00 240.5 F 11.6
14:00 238.7 F 10.0
16:00 252.4 F 16.9
18:00 264.3 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.8
1:00 272.7 F 29.5
2:00 272.1 F 28.8
3:00 271.9 F 29.1
4:00 272.6 F 29.5

Min 271.5 F 28.6
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.7 F 0.0
5:00 219.6 F 0.0
10:00 234.5 F 0.0
15:00 230.5 F 0.0
20:00 226.2 F 0.0
25:00 227.6 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:37
HEATING: 19:60
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:50
CYCLE COMPLETE

09/22/2022 01:58 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____