
Midmark M11 - v1.0.5
Total Cycles: 130

Name: RITTER2
Sterilizer ID: V2313228
06/23/2021 09:32 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.1 F 0.0
2:00 154.2 F 2.0
4:00 164.6 F 5.2
6:00 188.6 F 9.6
8:00 207.8 F 6.4
10:00 220.1 F 9.2
12:00 230.9 F 10.0
14:00 237.0 F 12.1
16:00 239.4 F 10.9
18:00 252.5 F 17.2
20:00 262.5 F 23.6
22:00 272.9 F 29.9

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 29.6
1:00 272.5 F 29.8
2:00 272.1 F 30.0
3:00 272.7 F 30.2
4:00 271.7 F 29.8

Min 271.6 F 29.5
Max 272.8 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.8 F 0.0
5:00 233.3 F 0.0
10:00 217.9 F 0.0
15:00 203.3 F 0.0
20:00 193.3 F 0.0
25:00 187.2 F 0.0
30:00 183.0 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 23:00
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:01:37
CYCLE COMPLETE

06/23/2021 10:33 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____