
Midmark M11 - v1.0.0
Total Cycles: 1487

Name: RITTER 1
Sterilizer ID: V2178198
12/29/2021 12:14 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 157.5 F 1.9
4:00 177.5 F 5.4
6:00 198.1 F 9.9
8:00 214.4 F 5.6
10:00 227.9 F 9.4
12:00 233.7 F 9.5
14:00 240.5 F 11.8
16:00 239.4 F 10.2
18:00 253.0 F 17.1
20:00 264.9 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.5
1:00 272.6 F 29.4
2:00 272.0 F 28.8
3:00 271.4 F 28.7
4:00 272.7 F 29.4

Min 271.3 F 28.4
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 225.6 F 0.0
10:00 231.4 F 0.0
15:00 227.6 F 0.0
20:00 229.2 F 0.0
25:00 227.0 F 0.0
30:00 227.2 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 21:51
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:16
CYCLE COMPLETE

12/29/2021 01:14 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____