
Midmark M11 - v1.0.0
Total Cycles: 2472

Name: RITTER 1
Sterilizer ID: V2178198
10/21/2022 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.4 F 0.0
2:00 185.4 F 4.6
4:00 195.5 F 9.4
6:00 213.7 F 7.3
8:00 229.9 F 8.8
10:00 237.9 F 10.4
12:00 241.1 F 11.7
14:00 239.2 F 10.1
16:00 252.9 F 17.0
18:00 264.7 F 24.0

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.6
1:00 272.1 F 29.2
2:00 272.5 F 29.0
3:00 271.5 F 28.6
4:00 272.7 F 29.4

Min 271.5 F 28.5
Max 272.8 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 234.0 F 0.0
10:00 227.1 F 0.0
15:00 231.0 F 0.0
20:00 227.7 F 0.0
25:00 227.8 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 19:44
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:58:19
CYCLE COMPLETE

10/21/2022 03:32 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____