

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3670

Name: RITTER2  
Sterilizer ID: V2313228  
06/05/2024 01:15 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 110.3 F 0.0  
2:00 111.2 F 0.8  
4:00 127.3 F 2.9  
6:00 155.3 F 6.1  
8:00 182.0 F 10.0  
10:00 204.5 F 7.0  
12:00 218.2 F 7.2  
14:00 228.2 F 11.8  
16:00 232.5 F 8.2  
18:00 242.2 F 12.7  
20:00 242.6 F 12.2  
22:00 254.4 F 18.3  
24:00 265.3 F 24.7  
26:00 270.9 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 272.7 F 29.7  
1:00 272.5 F 29.6  
2:00 272.3 F 29.6  
3:00 272.2 F 29.5  
4:00 272.2 F 29.6

Min 271.8 F 29.5  
Max 273.0 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 189.0 F 0.0  
5:00 230.7 F 0.0  
10:00 187.4 F 0.0  
15:00 180.7 F 0.0  
20:00 176.2 F 0.0  
25:00 174.4 F 0.0  
30:00 173.1 F 0.0  
Drying Complete

FILLING: 0:54  
HEATING: 26:29  
STERILIZING: 4:00  
VENTING: 3:22  
DRYING: 30:00  
TOTAL CYCLE: 01:04:45  
CYCLE COMPLETE

06/05/2024 02:19 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_