

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2574

Name: RITTER2  
Sterilizer ID: V2313228  
06/27/2023 11:33 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 147.2 F 0.0  
2:00 151.4 F 0.2  
4:00 173.2 F 3.9  
6:00 193.9 F 8.2  
8:00 214.2 F 13.4  
10:00 223.3 F 7.5  
12:00 235.3 F 12.3  
14:00 235.9 F 9.9  
16:00 237.3 F 9.4  
18:00 248.4 F 14.9  
20:00 260.0 F 21.5  
22:00 270.7 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.9  
1:00 271.7 F 29.3  
2:00 272.5 F 29.6  
3:00 271.8 F 29.1  
4:00 271.3 F 29.1

Min 271.2 F 28.8  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 199.0 F 0.0  
5:00 194.6 F 0.0  
10:00 202.4 F 0.0  
15:00 204.2 F 0.0  
20:00 202.7 F 0.0  
25:00 201.4 F 0.0  
30:00 198.2 F 0.0  
Drying Complete

FILLING: 0:52  
HEATING: 22:42  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 01:01:02  
CYCLE COMPLETE

06/27/2023 12:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_