
Midmark M11 - v1.0.0
Total Cycles: 1551

Name: RITTER 1
Sterilizer ID: V2178198
01/24/2022 07:46 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.6 F 0.0
2:00 76.7 F 0.0
4:00 101.6 F 0.3
6:00 117.6 F 2.4
8:00 152.2 F 5.4
10:00 177.3 F 8.9
12:00 191.3 F 12.5
14:00 208.6 F 3.9
16:00 217.2 F 7.3
18:00 228.0 F 11.0
20:00 229.2 F 6.8
22:00 240.1 F 11.7
24:00 237.2 F 9.3
26:00 250.4 F 15.6
28:00 261.4 F 21.8
30:00 271.6 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.3
3:00 272.5 F 29.3
4:00 271.9 F 28.6

Min 271.2 F 28.3
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 173.5 F 0.0
5:00 203.3 F 0.0
10:00 220.0 F 0.0
15:00 224.2 F 0.0
20:00 225.5 F 0.0
25:00 225.2 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 30:40
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 01:08:43
CYCLE COMPLETE

01/24/2022 08:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____