

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2021

Name: RITTER2  
Sterilizer ID: V2313228  
02/07/2023 04:10 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.8 F 0.0  
2:00 152.2 F 0.0  
4:00 180.8 F 2.9  
6:00 203.8 F 7.8  
8:00 222.6 F 12.7  
10:00 230.3 F 8.4  
12:00 239.9 F 11.4  
14:00 243.7 F 12.8  
16:00 243.8 F 12.4  
18:00 257.7 F 19.8  
20:00 269.8 F 27.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.7  
1:00 272.4 F 29.0  
2:00 272.4 F 29.5  
3:00 271.9 F 28.7  
4:00 272.7 F 29.5  
  
Min 271.5 F 28.6  
Max 272.8 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 182.4 F 0.0  
5:00 220.1 F 0.0  
10:00 227.3 F 0.0  
15:00 231.7 F 0.0  
20:00 227.5 F 0.0  
25:00 226.3 F 0.0  
30:00 226.3 F 0.0  
Drying Complete

FILLING: 0:57  
HEATING: 20:52  
STERILIZING: 4:00  
VENTING: 3:03  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

02/07/2023 05:08 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_