
Midmark M11 - v1.0.5
Total Cycles: 4008

Name: RITTER2
Sterilizer ID: V2313228
10/07/2024 03:38 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.2 F 0.0
2:00 145.6 F 2.2
4:00 157.8 F 5.4
6:00 184.1 F 9.4
8:00 202.6 F 8.6
10:00 219.6 F 7.7
12:00 232.0 F 12.5
14:00 233.4 F 9.7
16:00 235.7 F 9.5
18:00 246.6 F 14.5
20:00 258.2 F 20.9
22:00 268.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.5
1:00 272.8 F 30.3
2:00 271.9 F 29.7
3:00 271.5 F 29.8
4:00 272.7 F 30.3

Min 271.3 F 29.4
Max 272.9 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.4 F 0.0
5:00 190.5 F 0.0
10:00 193.3 F 0.0
15:00 189.5 F 0.0
20:00 185.2 F 0.0
25:00 182.1 F 0.0
30:00 179.8 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:10
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:02:02
CYCLE COMPLETE

10/07/2024 04:40 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____