
Midmark M11 - v1.0.0
Total Cycles: 3383

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2023 02:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 208.4 F 0.0
2:00 198.2 F 3.4
4:00 227.5 F 9.1
6:00 231.5 F 8.4
8:00 238.8 F 11.1
10:00 236.1 F 8.9
12:00 238.9 F 9.6
14:00 249.3 F 14.8
16:00 261.8 F 21.9
18:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.3
1:00 272.3 F 28.7
2:00 271.8 F 28.7
3:00 272.1 F 28.6
4:00 271.7 F 28.6

Min 271.2 F 28.1
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 233.2 F 0.0
10:00 225.5 F 0.0
15:00 230.6 F 0.0
20:00 224.3 F 0.0
25:00 218.7 F 0.0
30:00 204.1 F 0.0
Drying Complete

FILLING: 2:31
HEATING: 18:17
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:58:06
CYCLE COMPLETE

06/13/2023 03:05 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____